

# IMPROVE Carbon Analysis

Judith C. Chow (judith.chow@dri.edu)

John G. Watson

Dana L. Trimble

Desert Research Institute, Reno, NV

Presented at

the IMPROVE Steering Committee Meeting

Hot Springs, South Dakota

September 22, 2009

# Objectives

- Report progress on IMPROVE carbon analyses
- Present results of carbon analyzer quality control checks
- Enumerate improvements on carbon analyzers

# Summary of Carbon Lab Operations

- Acquired three additional Model 2001 carbon analyzers (Sept. 2008, and May 2009)
- Maintained 24 hours per day/7 days per week operation with 2 full-time and 2 part-time technologists, and 4 part-time students
- Recruited a full-time Electro-Mechanical Technician to perform maintenance on the analyzers (Sept. 2009)
- Analyzed over 23,000 IMPROVE samples

# **IMPROVE Samples for Carbon Analysis**

## **using the IMPROVE\_A Protocol**

**(7/08 – 6/09)**

<b>Sampling Period</b>	<b>Samples Received</b>	<b>Analysis Completion Date</b>
7/1/08-12/31/08	11,451	2/12/09
1/1/09-6/30/09	11,933	7/10/09

# Weekly Carbon Analyzer Quality Control Checks

(Effective 3/26/09)

	Sunday	Monday	Tuesday	Wed.	Thursday	Friday	Saturday
Morning (as part of startup)	Autocalib	Sucrose Injection*	Autocalib	Sucrose Injection	Autocalib	Sucrose Injection	Autocalib
Evening	CO <sub>2</sub> Injection	Autocalib	CO <sub>2</sub> Injection	Autocalib	CO <sub>2</sub> Injection	Autocalib	CO <sub>2</sub> Injection

\* Sucrose injection: 10 µl of a 5% solution  
TC ≈ 17.1 – 18.9 µg C/cm<sup>2</sup>


# Sucrose Calibration

(three times per week)

Analyzer ID

Target TC =  $18 \mu\text{g}/\text{cm}^2$  (5% acceptance range  $\approx 17.1$  to  $18.9 \mu\text{g}/\text{cm}^2$ )

◆ 6    ■ 7    ▲ 8    × 9    \* 10    ● 11    + 12    - 13    - 16    ◆ 18    ■ 19




# Daily AutoCalibration Procedure


- Inject CH<sub>4</sub> in He-only atmosphere (OC3)
- Inject CH<sub>4</sub> in He/O<sub>2</sub> atmosphere (EC1)
- Inject CH<sub>4</sub> in He/O<sub>2</sub> atmosphere for calibration peak area (CalSlope)

The three peaks should have similar area counts if the catalyst is in good condition and the calibration factor holds

# Temporal Variation AutoCalibration Chart for Analyzer #7




# Daily AutoCalibration Report (9/15/2009)



# Average CalPeak Area (CH<sub>4</sub>) Trend

Analyzer ID



# Carbon Analyzer Improvements

- Installed single stage regulators to ensure consistent ( $\pm 2 - 3\%$ ) flow pressure for He, He/O<sub>2</sub>, H<sub>2</sub>, and air
- Added aluminum-silica ceramic-fiber insulation around the FID to reduce baseline drift
- Applied NIST reference materials (RM8785, PM<sub>2.5</sub> fractions of SRM 1649a resuspended urban dust) for comparability between analyzers

